

Polystyrene Calibration Films

Part No.	Description	Price
0009-8157	Polystyrene calibration film for PhEur 2.2.24 calibration of IR and FTIR instruments, certified to 7 wave-numbers and certified as to absorption minima and maxima differentials at 2870 cm ⁻¹ and 2849.5cm ⁻¹ and at 1589 cm ⁻¹ and 1583 cm ⁻¹	\$174.00
0009-8158	Pair of PhEur 2.2.24 certified polystyrene calibration films, same as 0009-8157 but with 2 certified films	\$267.00
0009-8560	Polystyrene calibration film, certificate of traceability to seven (7) NIST frequencies (wave numbers) and certified as to absorption minima and maxima differentials at 2870cm ⁻¹ and 2849.5cm ⁻¹ and 1589cm ⁻¹ and 1583cm ⁻¹ as required by PhEur 2.2.24	\$202.00
0009-8448	Polystyrene calibration film, certificate of traceability to seven (7) NIST frequencies (wave numbers) and certified as to absorption minima and maxima differentials at 2870cm ⁻¹ and 2849.5cm ⁻¹ and 1589cm ⁻¹ and 1583cm ⁻¹ as required by PhEur 2.2.24 and also certified to PhJap 2.25	\$242.00
0009-8437	Pair of films same as 0009-8448	\$361.00
0009-9248	Custom Cut Polystyrene film cut to the size of any Horizontal ATR accessory (please specify, up to 2-3/4" square), certified to five (5) NIST 1921b frequencies	\$175.00

International Crystal Laboratories

11 Erie St., Garfield. NJ. 07642

www.internationalcrystal.net

iclmail@internationalcrystal.net

Phone: 973-478-8944 Fax: 973-478-4201

